

# Schedule

BS Tech Pte Ltd  
No. 41 Senoko Drive  
Singapore 758249

Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 1 of 34

FIELD OF TESTING : Calibration and Measurement

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>A.Electrical (Lab &amp; Site)</b></p> <p>1. Calibration of Resistance Measuring Instruments (2 Wire)</p> <p>0-10.9999 <math>\Omega</math>            11 <math>\Omega</math>-32.9999 <math>\Omega</math>            33 <math>\Omega</math>-109.9999 <math>\Omega</math>            110 <math>\Omega</math>-329.9999 <math>\Omega</math>            0.330 k<math>\Omega</math>-1.099999 k<math>\Omega</math>            1.1 k<math>\Omega</math>-3.299999 k<math>\Omega</math>            3.3 k<math>\Omega</math>-10.99999 k<math>\Omega</math>            11 k<math>\Omega</math>-32.99999 k<math>\Omega</math>            33 k<math>\Omega</math>-109.9999 k<math>\Omega</math>            110 k<math>\Omega</math>-329.9999 k<math>\Omega</math>            330k<math>\Omega</math>-1.0999999 M<math>\Omega</math>            1.1 M<math>\Omega</math>-3.299999 M<math>\Omega</math>            3.3 M<math>\Omega</math>-10.99999 M<math>\Omega</math>            11 M<math>\Omega</math>-32.99999 M<math>\Omega</math>            33 M<math>\Omega</math>-109.9999 M<math>\Omega</math>            110 M<math>\Omega</math>-329.9999 M<math>\Omega</math>            330 M<math>\Omega</math>-1100 M<math>\Omega</math></p>	<p>Technical Procedure Section BSE-01</p>	<p>0.012 <math>\Omega</math>            0.016 <math>\Omega</math>            0.022 <math>\Omega</math>            0.027 <math>\Omega</math>            0.000044 k<math>\Omega</math>            0.00027 k<math>\Omega</math>            0.00044 k<math>\Omega</math>            0.0019 k<math>\Omega</math>            0.0036 k<math>\Omega</math>            0.020 k<math>\Omega</math>            0.000041 M<math>\Omega</math>            0.0034 M<math>\Omega</math>            0.0014 M<math>\Omega</math>            0.015 M<math>\Omega</math>            0.060 M<math>\Omega</math>            1.6 M<math>\Omega</math>            14 M<math>\Omega</math></p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 2 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>2. Calibration of Resistance Measuring Instruments (4 wire)</p> <p>0 to 10.9999 <math>\Omega</math>            11 <math>\Omega</math> to 32.9999 <math>\Omega</math>            33 <math>\Omega</math> to 109.9999 <math>\Omega</math>            110 <math>\Omega</math> to 329.9999 <math>\Omega</math>            330 <math>\Omega</math> to 1.099999 k<math>\Omega</math>            1.1 k<math>\Omega</math> to 3.299999 k<math>\Omega</math>            3.3 k<math>\Omega</math> to 10.99999 k<math>\Omega</math>            11 k<math>\Omega</math> to 32.99999 k<math>\Omega</math>            33 k<math>\Omega</math> to 109.9999 k<math>\Omega</math></p>	<p>Technical Procedure Section BSE-01</p>	<p>0.012 <math>\Omega</math>            0.016 <math>\Omega</math>            0.018 <math>\Omega</math>            0.027 <math>\Omega</math>            0.000044 k<math>\Omega</math>            0.00027 k<math>\Omega</math>            0.00044 k<math>\Omega</math>            0.0019 k<math>\Omega</math>            0.0036 k<math>\Omega</math></p>
<p>3. Calibration of Resistance Measuring Instruments Wire (2 or 4 Wire)</p> <p>0.1 <math>\Omega</math> to 1 <math>\Omega</math>            1.1 <math>\Omega</math> to 10 <math>\Omega</math>            10.1 <math>\Omega</math> to 100 <math>\Omega</math>            100.1 <math>\Omega</math> to 1 k<math>\Omega</math>            1.1 k<math>\Omega</math> to 10 k<math>\Omega</math>            10.1 k<math>\Omega</math> to 100 k<math>\Omega</math>            100.1 k<math>\Omega</math> to 1 M<math>\Omega</math>            1.1 M<math>\Omega</math> to 10 M<math>\Omega</math>            10.1 M<math>\Omega</math> to 100 M<math>\Omega</math></p>	<p>Technical Procedure Section BSE-02</p>	<p>0.0024 <math>\Omega</math>            0.0035 <math>\Omega</math>            0.0088 <math>\Omega</math>            0.012 k<math>\Omega</math>            0.12 k<math>\Omega</math>            1.2 k<math>\Omega</math>            0.012 M<math>\Omega</math>            0.012 M<math>\Omega</math>            1.2 M<math>\Omega</math></p>
<p>4. Calibration of Milliohm meters</p> <p>10 m<math>\Omega</math>            30 m<math>\Omega</math>            50 m<math>\Omega</math>            100 m<math>\Omega</math>            330 m<math>\Omega</math>            470 m<math>\Omega</math></p>	<p>Technical Procedure Section BSE-02</p>	<p>1.2 m<math>\Omega</math>            3.5 m<math>\Omega</math>            5.8 m<math>\Omega</math>            12 m<math>\Omega</math>            38 m<math>\Omega</math>            54 m<math>\Omega</math></p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 3 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>5. High Voltage Resistance Measuring Instruments</p> <p>10 MΩ 50 MΩ 100 MΩ 200 MΩ 1 GΩ 2 GΩ 5 GΩ</p>	<p>Technical Procedure Section BSE-03</p>	<p>0.58 MΩ 2.9 MΩ 5.8 MΩ 12 MΩ 0.015 GΩ 0.12 GΩ 0.29 GΩ</p>
<p>6. Calibration of DC Voltage Measuring Instruments</p> <p>0 to 329.9999 mV 0.330 V to 3.299999 V 3.30 V to 32.99999 V 33 V to 329.9999 V 330 V to 1000.000 V</p>	<p>Technical Procedure Section BSE-04</p>	<p>0.0061 mV 0.000055 V 0.00036 V 0.0061 V 0.019 V</p>
<p>7. Calibration of DC Voltage Sourcing instruments</p> <p>0 to 100 mV 0.101 V to 1 V 1.1 V to 10 V 10.1 V to 100 V 100.1 V to 1000 V 1 kV to 10 kV 10.1 kV to 20 kV 20.1 kV to 30 kV 30.1 kV to 40 kV</p>	<p>Technical Procedure Section BSE-13</p>	<p>0.0076 mV 0.000047 V 0.0041 V 0.0053 V 0.0054 V 0.23 kV 0.47 kV 0.36 kV 0.93 kV</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 4 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
8. Calibration of AC Voltage Measuring Instruments	Technical Procedure Section BSE-05	
1.0 mV to 32.999 mV	10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 450 kHz	0.023 mV 0.0087 mV 0.0098 mV 0.028 mV 0.091 mV 0.23 mV
33 mV to 329.999 mV	10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 450 kHz	0.084 mV 0.044 mV 0.048 mV 0.084 mV 0.23 mV 0.57 mV
0.330 V to 3.2999 V	10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 450 kHz	0.00081 V 0.00043 V 0.00054 V 0.00081 V 0.0019 V 0.0067 V
3.3 V to 32.9999 V	10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 50 kHz 50 kHz to 90 kHz	0.0082 V 0.0044 V 0.0066 V 0.024 V
33 V to 329.999 V	45 Hz to 1 kHz 1 kHz to 10 kHz	0.027 V 0.057 V
33 V to 200 V	10 kHz to 20 kHz 20 kHz to 50 kHz 50 kHz to 90 kHz	0.074 V 0.086 V 0.35 V
330 V to 750 V	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 8kHz	0.18 V 0.12 V 0.18 V

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 5 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
9. Calibration of AC Voltage Sourcing Instruments	Technical Procedure Section BSE-14	
1 mV to 100 mV	10 Hz to 1 kHz 1 kHz to 50 kHz	0.086 mV 0.15 mV
0.101 V to 1 V	10 Hz to 1 kHz 1 kHz to 50 kHz 1 kHz to 100 kHz	0.00068 V 0.00085 V 0.0070 V
1.1 V to 10 V	10 Hz to 1 kHz 1 kHz to 50 kHz	0.0068 V 0.015 V
10.1 V to 100 V	10 Hz to 1 kHz 1 kHz to 50 kHz	0.068 V 0.15 V
100.1 V to 700 V	10 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 50 kHz	0.49 V 0.66 V 1.1 V
0.701 kV to 10 kV 10.1 kV to 20 kV 20.1 kV to 28 kV	60 Hz 60 Hz 60 Hz	0.61 kV 1.2 kV 1.7 kV
10. Calibration of DC Current Measuring Instruments	Technical Procedure Section BSE-06	
1 µA to 329.999 µA 0.330 mA to 3.29999 mA 3.30 mA to 32.9999 mA 33 mA to 329.999 mA 0.330 A to 1.0 A 1.01 A to 2.0 A 2.01 A to 10.0 A 10.01 A to 20.5 A		0.054 µA 0.00030 mA 0.0028 mA 0.028 mA 0.00020 A 0.0012 A 0.0047 A 0.019 A

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 6 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>11. Calibration of DC Current Clamp Meters</p> <p>0 to 10 A 10.1 A to 50 A 50.1 A to 100 A 100.1 A to 500 A 501 A to 1000 A</p>	<p>Technical Procedure Section BSE-07</p>	<p>0.1 A 0.38 A 0.54 A 2.5 A 4.7 A</p>
<p>12. Calibration of DC Current Sourcing Instruments</p> <p>10 <math>\mu</math>A to 100 <math>\mu</math>A 0.101 mA to 1 mA 1.1 mA to 10 mA 10.1 mA to 100 mA 0.101 A to 1.0 A 1.01 A to 2 A</p>	<p>Technical Procedure Section BSE-15</p>	<p>0.42 <math>\mu</math>A 0.0024 mA 0.0062 mA 0.058 mA 0.097 A 0.0029 A</p>
<p>13. Calibration of AC Current Measuring instruments</p> <p>29 <math>\mu</math>A to 329.999 <math>\mu</math>A</p> <p>0.330 mA to 3.2999 mA</p> <p>3.3 mA to 32.999 mA</p> <p>33 mA to 329.99 mA</p> <p>0.330 A to 1.0 A</p>	<p>Technical Procedure Section BSE-08</p> <p>10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 30 kHz</p> <p>10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 30 kHz</p> <p>10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 30 kHz</p> <p>10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 30 kHz</p> <p>10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz</p>	<p>0.46 <math>\mu</math>A 0.40 <math>\mu</math>A 4.4 <math>\mu</math>A</p> <p>0.0033 mA 0.0015 mA 0.026 mA</p> <p>0.025 mA 0.012 mA 0.11 mA</p> <p>0.25 mA 0.13 mA 1.2 mA</p> <p>0.0015 A 0.00052 A 0.023 A</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 7 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
1.01 A to 2.0 A	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.0023 A 0.0013 A 0.043 A
2.01 A to 10.0 A	45 Hz to 1 kHz 1 kHz to 5 kHz	0.0013 A 0.0013 A
10.1 A to 20.5 A	45 Hz to 1 kHz 1 kHz to 5 kHz	0.032 A 0.47 A
14. Calibration of AC Current Clamp meters	Technical Procedure Section BSE-08	
0.1 A to 5A 5.1 A to 10 A 10.1 A to 50 A 50.1 A to 100 A 100.1 A to 500 A 500.1 A to 1000 A	30 Hz to 50 Hz	0.073 A 0.1 A 0.38 A 0.54 A 2.5 A 4.7 A
15. Calibration of AC Current Sourcing Instruments	Technical Procedure Section BSE-16	
10 mA to 100 mA 0.101 A to 1 A 1.01 A to 2 A	10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz	0.35 mA 0.0013 A 0.0041A
16. Calibration of Frequency Measuring Instruments	Technical Procedure Section BSE-09	
0.01 Hz to 1.199 kHz 1.200 kHz to 100 kHz 100.1 kHz to 2.000 MHz		0.00029 Hz 0.00025 kHz 0.000004 MHz

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 8 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<b>17. Calibration of Capacitance Measuring Instruments</b>  0.19 nF to 0.3999 nF 0.4 nF to 1.0999 nF 1.1 nF to 10.9999 nF 11 nF to 32.9999 nF 33 nF to 109.999 nF 110 nF to 329.999 nF 0.33 µF to 1.09999 µF 1.1µF to 3.29999 µF 3.3 µF to 10.9999 µF 11µF to 32.9999 µF 33 µF to 109.9999 µF 110 µF to 329.999 µf 0.33 mF to 1.09999 mF 1.1 mF to 3.2999 mF	Technical Procedure Section BSE-10  1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 100 Hz 100 Hz 100 Hz 100 Hz 50 Hz 50 Hz 50 Hz 20 Hz	0.0094 nF 0.012 nF 0.032 nF 0.15 nF 0.31 nF 0.89 nF 0.0030 µF 0.0094 µF 0.031 µF 0.13 µF 0.47 µF 1.4 µF 0.0047 mF 0.014 mF
<b>18. Calibration of Inductance Measuring Instruments</b>  1 mH 10 mH 20 mH 30 mH 50 mH 100 mH 1 H 10 H	Technical Procedure Section BSE-11  1kHz	0.012 mH 0.066 mH 0.12 mH 0.18 mH 0.31 mH 0.61 mH 0.0062 H 0.069 H
<b>19. Calibration of Electrical Power meter Instruments</b>  0.1 W to 1000 W 1.1 kW to 2 kW 2.1 kW to 14 kW	Technical Procedure Section BSE-17  40 Hz to 60 Hz	0.0081 W 0.0018 kW 0.0087 kW
<b>20. Calibration of Digital Timer, Analog Timer, Stop Watch</b>  600 s to 1800 s 1801 s to 3600 s	Technical Procedure Section BSE-12	0.61 s 0.85 s



# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 9 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p><b>B. Mechanical (Lab &amp; Site)</b></p> <p>1. Absolute Pressure Instruments</p> <p>Type of Instruments:                      Absolute Gauge                      Absolute Recorder                      Absolute Calibrator                      Digital Indicator                      Transmitters                      Transducers                      Barometer                      Manometer</p> <p>Range:                      0 to 1600 mbar abs</p> <p>2. Non-Oil/Oxygen Pressure Instruments</p> <p>Range :                      0.000 to 2.000 "H<sub>2</sub>O                      -1000 mbar to 30 psi                      0 to 1000 psi                      10 psi to 10000 psi</p> <p>3. Pneumatic Instruments</p> <p>Type of Instruments :</p> <p>Pneumatic Gauge                      Calibrators                      Digital Indicators                      Data Logger                      Manometers                      Transducer                      Chart Recorders                      Magnehelic gauge</p> <p>Range :                      0.000 to 2.000 "H<sub>2</sub>O                      -1000 mbar to 30 psi                      0 to 1000 psi</p>	<p>Technical Procedure                      Section BSM-06</p> <p>Technical Procedure                      Section                      BSM-01                      BSM-02                      BSM-03,&amp;                      BSM-04</p> <p>Technical Procedure                      Section BSM-02                      BSM-03 &amp;                      BSM-04</p>	<p>7.3 mbar</p> <p>0.009 "H<sub>2</sub>O                      0.14%                      0.6 psi                      9.3 psi</p> <p>0.009 "H<sub>2</sub>O                      0.14%                      0.6 psi</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 10 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>4. Vacuum Instruments</p> <p>Type of Instruments :</p> <p>Digital Indicator Torr Meter Vacuum Gauges Digital Indicator Manometer Calibrators Magnehelic Gauge and High Vacuum Sensors</p> <p>Range :</p> <p>10 to 1 Torr 1 Torr to 1E to 4 Torr 1E-4 Torr to 1E-5 Torr 1E-5 Torr to 1E-6 Torr</p>	<p>Technical Procedure Section BSM-06</p>	<p>5.3 % 27 % 25 % 29 %</p>
<p>5. Hydraulic Pressure Instruments</p> <p>Type of Instruments :</p> <p>Oil Pressure Gauge, Distilled Water Pressure Gauge, Calibrator, Digital Indicator, Manometers, Controller, Data Logger, Chart Recorder, Pressure Switches, Pressure transducer and Transmitters</p> <p>Range :</p> <p>0 to 16 000 psi 0 to 2500 bar</p>	<p>Technical Procedure Section BSM-01</p>	<p>0.022 % of Reading 6.4 bar</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 11 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>6. Dead Weight Testers</p> <p>Type of Instruments :</p> <p>Hydraulic Oil Dead Weight Tester (Cross Float)</p> <p>Range : 0 to 16000 psi</p>	<p>Technical Procedure Section BSM-01</p>	<p>0.022 % of Reading</p>
<p>7. Differential Pressure</p> <p>Type of Instruments :</p> <p>Pneumatic Differential Gauge Switches Pneumatic Differential Indicators and Magnehelic Gauge Manometers Flowmeters Anemometer Velometer Balometer</p> <p>Range : 0 to 2 "H<sub>2</sub>O</p>	<p>Technical Procedure Section BSM-04</p>	<p>0.009 "H<sub>2</sub>O</p>
<p>8. Force Gauge Instruments</p> <p>Range : 0 to 1000 psi 0 to 16000 psi</p>	<p>Technical Procedure Section BSM-01 &amp; BSM-02</p>	<p>0.6 psi 0.022 % of reading</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 12 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>9. Pressure Relief Valve Instruments</p> <p>Range : Gas Up to 70 bar</p>	<p>Technical Procedure Section BSM-09</p>	<p>0.53 % of F.S</p>
<p>10. Non Contact &amp; Contact Tachometer Instruments</p> <p>Non Contact Tachometer Range : 10 to 99 999 rpm</p> <p>Contact Tachometer Range : 0 to 4800 rpm</p>	<p>Technical Procedure Section BSM-08</p>	<p>3.6 rpm  2.5 rpm</p>
<p>11. Calibration of Balances and Weighing Scale</p> <p>Type of Instruments : Weighing Scales Compact Scales Industrial Balances Electronics Balances Environmental Balances Dust &amp; Waterproof Digital Scales Electronics Pocket Balances Moisture Balance Top Loading Balances Bench / Counting Scale Floor Scale Platform Scale</p>	<p>Technical Procedure Section BSM-07</p>	

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 13 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>1 mg to 20 mg 20 mg to 50 mg 50 mg to 100 mg 100 mg to 200 mg 200 mg to 500 mg 500 mg to 500 g 500 g to 5000 g 5000 g to 10000 g 10000 g to 24000 g 24 kg to 60 kg 60 kg to 200 kg 200 kg to 300 kg 300 kg to 500 kg 500 kg to 600 kg</p> <p>12. Calibration of Standard Weights Test Weights Precision Weights</p> <p>OIML Class M3 Weights or equivalent</p> <p>1 g to 50 g 50 g to 200 g 200 g to 500 g 500 g to 1 kg 1 kg to 2 kg 2 kg to 5 kg 5 kg to 10 kg 10 kg to 20 kg</p> <p>OIML Class M2 Weights or equivalent</p> <p>1 kg to 5 kg 5 kg to 10 kg 10 kg to 20 kg</p>	<p>Technical Procedure Section BSM-12</p>	<p>0.14 mg 0.14 mg 0.14 mg 0.14 mg 0.14 mg 0.4 mg 3 mg 20 mg 100 mg 0.011 kg 0.02 kg 0.05 kg 0.13 kg 0.2 kg</p> <p>0.0002 g 0.0002 g 0.012 g 0.012 g 0.012 g 0.012 g 0.058 kg 0.12 kg</p> <p>0.0058 kg 0.029 kg 0.058 kg</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 14 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>OIML Class M1 Weights or equivalent</p> <p>2 kg to 5 kg 2 kg to 10 kg 10 kg to 20 kg</p> <p>13. Torque Measuring Instruments</p> <p>Type of Instruments : Torque Hand Tools Round dial gauge Torque Screwdriver Torque Gauge (Digital / Analog) Dial Indicating Type Torque Wrench Indicator Plate Torque Wrench Adjustable Torque Wrench Rotating Torque Screw Driver Preset Torque Wrench Data Printout Digital Torque Wrench Hydraulic Torque Gauge</p> <p>Range : 0 to 1.6 N.m 1.6 to 10 N.m 10 to 100 N.m 100 to 500 N.m 500 to 1000 N.m 1000 to 1500 N.m 1500 to 2000 N.m 2000 to 2500 N.m 2500 to 2700 N.m</p>	<p>Technical Procedure Section BSM-10</p>	<p>0.012 kg 0.029 kg 0.058 kg</p> <p>0.08 N.m 0.32 N.m 0.89 N.m 6.5 N.m 13.0 N.m 15.0N.m 16.0 N.m 27.0 N.m 28.0 N.m</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 15 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>14. Torque Meter Tester</p> <p>Type of Instruments : Torque Checker Torque Meter Tester Torque Analyzer</p> <p>Range : Up to 100 kgf.cm</p>	<p>Technical Procedure Section BSM-11</p>	<p>0.15 kgf.cm</p>
<p>15. Force Measuring Instruments</p> <p>Type of Instruments : Push Pull Gauge Force Gauge Load Cell Tension Meter Gram Gauge Manual Test Stands Motorized Test Stands Dial Tension Gauge Tension Gauge Wire Tension Meter Load Gauge Spring Balance Tensile Tester Cable Tension Meter Force Indicator Dynamometer Belt Tension Gauge Strap Tension Meter</p> <p>Range : 1 gf to 100 gf 100 gf to 1kgf 1 kgf to 100 kgf 100 kgf to 400 kgf 400 kgf to 600 kgf</p>	<p>Technical Procedure Section BSM-13</p>	<p>0.005 kgf.cm 0.04 kgf.cm 0.04 kgf.cm 0.25 kgf.cm 0.37 kgf.cm</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 16 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)																																	
<p>16. Rubber Hardness Tester</p> <p>Type of Instruments            Hand Durometer            Digital Durometer            Pencil Durometer            Multi Scale Durometer            Dial Durometer            Tire Hardness Durometer            Pocket Durometer</p> <p>Range :            Up to 100 Div            Durometer Type A, B, C</p>	<p>Technical Procedure            Section BSM-14</p> <p>Type A, B, C, D</p>	<p>0.3 Div</p>																																	
<p><b>C. Temperature (Lab &amp; Site)</b></p> <p>1. Temperature Measuring Instruments</p> <p>a. Glass Thermometer (Lab)</p> <table border="0"> <tr> <td>Range:</td> <td>Scale Div</td> <td></td> </tr> <tr> <td>- 75 ~ -30 °C</td> <td>0.1 °C</td> <td>0.20 °C</td> </tr> <tr> <td>- 30 ~ 30 °C</td> <td>0.1 °C</td> <td>0.17 °C</td> </tr> <tr> <td>30 ~ 100 °C</td> <td>0.1 °C</td> <td>0.10 °C</td> </tr> <tr> <td>100 ~ 300 °C</td> <td>0.1 °C</td> <td>0.10 °C</td> </tr> </table> <p>b. RTD Sensor/ Thermister Probe without indicator</p> <table border="0"> <tr> <td>Range :</td> <td></td> <td><u>Lab</u> / <u>Site</u></td> </tr> <tr> <td>- 75 ~ -20 °C</td> <td></td> <td>0.19 °C / -</td> </tr> <tr> <td>- 20 ~ 30 °C</td> <td></td> <td>0.16 °C / 0.22 °C</td> </tr> <tr> <td>30 ~ 100 °C</td> <td></td> <td>0.06 °C / 0.22 °C</td> </tr> <tr> <td>100 ~ 300 °C</td> <td></td> <td>0.07 °C / 0.83 °C</td> </tr> <tr> <td>300 ~ 500 °C</td> <td></td> <td>0.22 °C / 0.83 °C</td> </tr> </table>	Range:	Scale Div		- 75 ~ -30 °C	0.1 °C	0.20 °C	- 30 ~ 30 °C	0.1 °C	0.17 °C	30 ~ 100 °C	0.1 °C	0.10 °C	100 ~ 300 °C	0.1 °C	0.10 °C	Range :		<u>Lab</u> / <u>Site</u>	- 75 ~ -20 °C		0.19 °C / -	- 20 ~ 30 °C		0.16 °C / 0.22 °C	30 ~ 100 °C		0.06 °C / 0.22 °C	100 ~ 300 °C		0.07 °C / 0.83 °C	300 ~ 500 °C		0.22 °C / 0.83 °C	<p>Technical Procedure            Section BST 01-H            BST 01-L</p> <p>Technical Procedure Section            BST 01-H, BST 01-L</p>	<p>0.20 °C            0.17 °C            0.10 °C            0.10 °C</p> <p><u>Lab</u> / <u>Site</u>            0.19 °C / -            0.16 °C / 0.22 °C            0.06 °C / 0.22 °C            0.07 °C / 0.83 °C            0.22 °C / 0.83 °C</p>
Range:	Scale Div																																		
- 75 ~ -30 °C	0.1 °C	0.20 °C																																	
- 30 ~ 30 °C	0.1 °C	0.17 °C																																	
30 ~ 100 °C	0.1 °C	0.10 °C																																	
100 ~ 300 °C	0.1 °C	0.10 °C																																	
Range :		<u>Lab</u> / <u>Site</u>																																	
- 75 ~ -20 °C		0.19 °C / -																																	
- 20 ~ 30 °C		0.16 °C / 0.22 °C																																	
30 ~ 100 °C		0.06 °C / 0.22 °C																																	
100 ~ 300 °C		0.07 °C / 0.83 °C																																	
300 ~ 500 °C		0.22 °C / 0.83 °C																																	



# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 17 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>c. RTD Sensor/ Thermister Probe with Indicator</p> <p>Range :</p> <p>-75 ~ -20 °C</p> <p>-20 ~ 30 °C</p> <p>30 ~ 100 °C</p> <p>100 ~ 300 °C</p> <p>300 ~ 500 °C</p> <p>d. Digital Thermometer with sensor, Temperature Switch, Temperature Transmitter, Temperature Transducer, Data Logger with Sensor, Temperature Controller with Sensor Temperature Indicator with Sensor Sensor, Temperature Recorder with Sensor</p> <p>Range :</p> <p>-75 ~ -20 °C</p> <p>-20 ~ 30 °C</p> <p>30 ~ 100 °C</p> <p>100 ~ 300 °C</p> <p>300 ~ 500 °C</p> <p>500 ~ 1000 °C</p> <p>e. Thermocouple without Indicator</p> <p>Range :</p> <p>-75 ~ -20 °C</p> <p>-20 ~ 30 °C</p> <p>30 ~ 100 °C</p> <p>100 ~ 300 °C</p> <p>300 ~ 500 °C</p> <p>500 ~ 1000 °C</p>	<p>Technical Procedure Section BST 01-H, BST 01-L &amp; BST-02</p>	<p><u>Lab</u> / <u>Site</u></p> <p>0.2 °C / -</p> <p>0.2 °C / 0.2 °C</p> <p>0.1 °C / 0.2 °C</p> <p>0.1 °C / 0.8 °C</p> <p>0.2 °C / 0.8 °C</p> <p><u>Lab</u> / <u>Site</u></p> <p>0.20 °C / -</p> <p>0.17 °C / 0.23 °C</p> <p>0.09 °C / 0.23 °C</p> <p>0.19 °C / 0.84 °C</p> <p>0.23 °C / 0.83 °C</p> <p>1.5 °C / 1.5 °C</p> <p><u>Lab</u> / <u>Site</u></p> <p>0.27 °C / -</p> <p>0.27 °C / 0.32 °C</p> <p>0.21 °C / 0.30 °C</p> <p>0.53 °C / 1.0 °C</p> <p>0.61 °C / 1.0 °C</p> <p>1.6 °C / 1.6 °C</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 18 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>f. Thermocouple with Indicator</p> <p>Range :</p> <p>-75 ~ -20 °C</p> <p>-20 ~ 30 °C</p> <p>30 ~ 100 °C</p> <p>100 ~ 300 °C</p> <p>300 ~ 500 °C</p> <p>500 ~ 1000°C</p> <p>2. Dry Block Calibrator (Lab/Site)</p> <p>Range :</p> <p>-40 ~ 650 °C</p> <p>3. Ice Point</p> <p>Range :</p> <p>0 °C</p> <p>4. (a) Calibration of Temperature Indicator (Lab/Site)</p> <p>Type J :</p> <p>-210 ~ 150 °C</p> <p>150 ~ 1200 °C</p> <p>Type K :</p> <p>-200 ~ 200 °C</p> <p>200 ~ 1300 °C</p> <p>Type T :</p> <p>-250 ~ -150 °C</p> <p>-150 ~ 400 °C</p> <p>Type R :</p> <p>0 ~ 250 °C</p> <p>250 ~ 500 °C</p> <p>500 ~ 1750 °C</p>	<p>Technical Procedure Section BST 01-H, BST 01-L &amp; BST-02</p> <p>Technical Procedure Section BST 04</p> <p>Technical Procedure Section BST 05</p>	<p><u>Lab</u> / <u>Site</u></p> <p>0.22 °C / -</p> <p>0.22 °C / 0.28 °C</p> <p>0.13 °C / 0.25 °C</p> <p>0.15 °C / 1.0 °C</p> <p>0.32 °C / 1.0 °C</p> <p>1.6 °C / 1.6 °C</p> <p>1.2 °C</p> <p>0.054 °C</p> <p>0.35 °C</p> <p>0.31 °C</p> <p>0.52 °C</p> <p>0.57 °C</p> <p>0.78 °C</p> <p>0.32 °C</p> <p>0.91 °C</p> <p>0.75 °C</p> <p>0.77 °C</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 19 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>Type S : 0 ~ 1200 °C 1200 ~ 1750 °C</p> <p>Type B : 600 ~ 1000 °C 1000 ~ 1800 °C</p> <p>Type N : - 200 ~ 50 °C 50 ~ 1300 °C</p> <p>Type E : -100 ~ 1000 °C</p> <p>4. (b) Calibration of Thermocouple Sourcing Instrument(Lab/Site)</p> <p>Type J : -210 ~ 150 °C 150 ~ 1200 °C</p> <p>Type K : -200 ~ 200 °C 200 ~ 1300 °C</p> <p>Type T : -250 ~ -150 °C -150 ~ 400 °C</p> <p>Type R : 0 ~ 250 °C 250 ~ 500 °C 500 ~ 1750 °C</p> <p>Type S : 0 ~ 1200 °C 1200 ~ 1750 °C</p> <p>Type B : 600 ~ 1000 °C 1000 ~ 1800 °C</p>	<p>Technical Procedure Section BST 05</p>	<p>0.83 °C 0.80 °C</p> <p>0.83 °C 0.70 °C</p> <p>0.96 °C 0.90 °C</p> <p>0.35 °C</p> <p>0.36 °C 0.32 °C</p> <p>0.53 °C 0.57 °C</p> <p>0.80 °C 0.33 °C</p> <p>0.92 °C 0.79 °C 0.77 °C</p> <p>0.89 °C 0.81 °C</p> <p>0.87 °C 0.70 °C</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 20 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>Type N : -200 ~ 260 °C 260 ~ 1300 °C</p> <p>Type E : -100 ~ 1000 °C</p> <p>5. (a) PRT Electrical Simulation Instruments (Lab/Site)</p> <p>Range : Type PT 100 (385) -100 ~ 800 °C</p> <p>(b) PRT Sourcing Instruments (Lab/Site)</p> <p>Range : Type PT 100 (385) -100 ~ 800 °C</p> <p>6. Controlled Temperature Enclosure (Lab/Site)</p> <p>a. Freezers Oven &amp; Furnaces</p> <p>Range : -75 ~ 250 °C 250 ~ 500 °C 500 ~ 800 °C 800 ~ 1000 °C</p> <p>b. Autoclaves</p> <p>7. Humidity Chamber(Lab/Site)</p> <p>Humidity Range : 10 ~ 95 % rh.( 5~50 °C)</p> <p>Temperature Range : 5 ~ 50 °C (10 ~ 95 %) rh.</p>	<p>Technical Procedure Section BST 05</p> <p>Technical Procedure Section BST 06</p> <p>Technical Procedure Section BST 06</p>	<p>0.98 °C 0.90 °C</p> <p>0.36 °C</p> <p>0.27 °C</p> <p>0.20 °C</p> <p>2.5 °C 2.8 °C 3.9 °C 4.2 °C</p> <p>1.4 °C</p> <p>2.9 %rh. 1.4 °C</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 21 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)
<p>8. Calibration of Surface Probes with/without Indicator (Lab/Site)</p> <p>Range : 50 ~ 250 °C 250 ~ 450 °C</p> <p>9. Humidity / Temperature /Dew Point Measuring Temperature Instruments (Lab/Site)</p> <p>Humidity : 10 ~ 95 %rh ( 5 ~ 50 °C)</p> <p>Temperature : 5 ~ 23 °C ( 10 ~ 95 %rh) 23 ~ 50 °C ( 10 ~ 95 %rh)</p>	<p>Technical Procedure Section BST 08</p>	<p>3.0 °C 5.5 °C</p> <p>3.0 % rh</p> <p>0.5 °C 0.5 °C</p>
<p><b>D.Dimensional (Lab &amp; Site)</b></p> <p>1. Micrometer Measuring Instruments</p> <ul style="list-style-type: none"> <li>a. External Micrometer</li> <li>b. Interchangeable Micrometer</li> <li>c. Digimatic Micrometer</li> <li>d. Outside Micrometer</li> <li>e. Caliper Type Micrometer</li> <li>f. Screw Thread Micrometer</li> <li>g. Gear tooth Micrometer</li> <li>h. Disk Micrometer</li> <li>i. Sheet Metal Micrometer</li> <li>j. Spline Micrometer</li> <li>k. Quick-Mini</li> <li>l. Tube Micrometer</li> <li>m. Point Micrometer</li> <li>n. V-Anvil Micrometer</li> <li>o. Blade Micrometer</li> <li>p. Groove Micrometer</li> <li>q. Can Seam Micrometer</li> </ul>	<p>Technical Procedure Section BSD-01</p>	

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 22 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)																												
<p>r. Digit outside Micrometer s. Digimatic Bench Micrometer t. Dial Snap Meter u. Indicating Micrometer</p> <p>Range:</p> <p>Flatness</p> <p>Parallelism: 0 mm to 500 mm 501 mm to 1000 mm 1001 mm to 1500 mm</p> <p>Accuracy: 0 mm to 500 mm 501 mm to 1000 mm 1001 mm to 1500 mm</p> <p>2. End Rod measurements</p> <p>Standard Rod / Extension road</p> <p>Range: 0 mm to 500 mm 501 mm to 1000 mm 1001 mm to 1500 mm</p> <p>3. Caliper Measuring Instruments</p> <p>a. Super Caliper b. Solar Caliper c. Coolant Proof Caliper d. Digimatic Caliper e. Vernier Caliper f. Dial Caliper g. Carbon Fiber Caliper h. Long jaw Caliper i. Back-jaw Caliper j. Blade type Caliper</p>	<p>Technical Procedure Section BSD-02</p> <p>Technical Procedure Section BSD-03</p>	<table border="0"> <tr> <td><u>Lab</u></td> <td><u>On-site</u></td> </tr> <tr> <td>1 µm</td> <td>2 µm</td> </tr> <tr> <td><u>Lab</u></td> <td><u>On-site</u></td> </tr> <tr> <td>2 µm</td> <td>3 µm</td> </tr> <tr> <td>2 µm</td> <td>3 µm</td> </tr> <tr> <td>3 µm</td> <td>3 µm</td> </tr> <tr> <td><u>Lab</u></td> <td><u>On-site</u></td> </tr> <tr> <td>1 µm</td> <td>2 µm</td> </tr> <tr> <td>2 µm</td> <td>2 µm</td> </tr> <tr> <td>3 µm</td> <td>3 µm</td> </tr> <tr> <td><u>Lab</u></td> <td><u>On-site</u></td> </tr> <tr> <td>2 µm</td> <td>3 µm</td> </tr> <tr> <td>3 µm</td> <td>4 µm</td> </tr> <tr> <td>3 µm</td> <td>4 µm</td> </tr> </table>	<u>Lab</u>	<u>On-site</u>	1 µm	2 µm	<u>Lab</u>	<u>On-site</u>	2 µm	3 µm	2 µm	3 µm	3 µm	3 µm	<u>Lab</u>	<u>On-site</u>	1 µm	2 µm	2 µm	2 µm	3 µm	3 µm	<u>Lab</u>	<u>On-site</u>	2 µm	3 µm	3 µm	4 µm	3 µm	4 µm
<u>Lab</u>	<u>On-site</u>																													
1 µm	2 µm																													
<u>Lab</u>	<u>On-site</u>																													
2 µm	3 µm																													
2 µm	3 µm																													
3 µm	3 µm																													
<u>Lab</u>	<u>On-site</u>																													
1 µm	2 µm																													
2 µm	2 µm																													
3 µm	3 µm																													
<u>Lab</u>	<u>On-site</u>																													
2 µm	3 µm																													
3 µm	4 µm																													
3 µm	4 µm																													



# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 24 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)	
0 mm to 1000 mm    10 µm 20 µm 50 µm  0 mm to 1500 mm    10 µm 20 µm 50 µm		10 µm 20 µm 50 µm	10 µm 20 µm 50 µm
<b>5. Depth Measuring Instruments</b>  a. Vernier Depth Gage b. Digimatic Depth Gage c. Dial Depth Gage d. Depth Gage e. Tire Thread Depth Gage f. Pitch Depth Gage g. Thread Depth Gage  Range:                    Resolution: 0 mm to 500 mm    10 µm 20 µm 50 µm  Range:                    Resolution: 0 mm to 1000 mm    10 µm 20 µm 50 µm	Technical Procedure Section BSD-05	<u>Lab</u> 10 µm 20 µm 50 µm  <u>Lab</u> 10 µm 20 µm 50 µm	<u>On-site</u> 10 µm 20 µm 50 µm  <u>On-site</u> 10 µm 20 µm 50 µm
<b>6. Internal Measuring Instruments</b>  a. Stick Micrometers b. Digimatic Tubular Micrometers c. Tubular Micrometers d. Inside Micrometer	Technical Procedure Section BSD-06		



# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 25 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)												
<p>Accuracy:</p> <p>Range : 0 mm to 1500 mm</p> <p>Extension Rods:</p> <p>Range: 0 mm to 500 mm 501 mm to 1500 mm</p> <p>7. Depth Micrometers</p> <p>Flatness :</p> <p>Parallelism: 0 mm to 300 mm</p> <p>Accuracy: 0 mm to 300 mm</p> <p>Extension Rods: 0 mm to 300 mm</p>	<p>Technical Procedure Section BSD-07</p>	<table border="0"> <tr> <td><u>Lab</u> 3 <math>\mu</math>m</td> <td><u>On-site</u> 4 <math>\mu</math>m</td> </tr> <tr> <td><u>Lab</u> 2 <math>\mu</math>m 3 <math>\mu</math>m</td> <td><u>On-site</u> 3 <math>\mu</math>m 4 <math>\mu</math>m</td> </tr> <tr> <td><u>Lab</u> 2 <math>\mu</math>m</td> <td><u>On-site</u> 3 <math>\mu</math>m</td> </tr> <tr> <td><u>Lab</u> 2 <math>\mu</math>m</td> <td><u>On-site</u> 3 <math>\mu</math>m</td> </tr> <tr> <td><u>Lab</u> 1 <math>\mu</math>m</td> <td><u>On-site</u> 2 <math>\mu</math>m</td> </tr> <tr> <td><u>Lab</u> 2 <math>\mu</math>m</td> <td><u>On-site</u> 3 <math>\mu</math>m</td> </tr> </table>	<u>Lab</u> 3 $\mu$ m	<u>On-site</u> 4 $\mu$ m	<u>Lab</u> 2 $\mu$ m 3 $\mu$ m	<u>On-site</u> 3 $\mu$ m 4 $\mu$ m	<u>Lab</u> 2 $\mu$ m	<u>On-site</u> 3 $\mu$ m	<u>Lab</u> 2 $\mu$ m	<u>On-site</u> 3 $\mu$ m	<u>Lab</u> 1 $\mu$ m	<u>On-site</u> 2 $\mu$ m	<u>Lab</u> 2 $\mu$ m	<u>On-site</u> 3 $\mu$ m
<u>Lab</u> 3 $\mu$ m	<u>On-site</u> 4 $\mu$ m													
<u>Lab</u> 2 $\mu$ m 3 $\mu$ m	<u>On-site</u> 3 $\mu$ m 4 $\mu$ m													
<u>Lab</u> 2 $\mu$ m	<u>On-site</u> 3 $\mu$ m													
<u>Lab</u> 2 $\mu$ m	<u>On-site</u> 3 $\mu$ m													
<u>Lab</u> 1 $\mu$ m	<u>On-site</u> 2 $\mu$ m													
<u>Lab</u> 2 $\mu$ m	<u>On-site</u> 3 $\mu$ m													
<p>8. Plunger Indicating Instruments</p> <ul style="list-style-type: none"> <li>a. Dial indicator</li> <li>b. Digimatic indicator</li> <li>c. Hicator</li> <li>d. Signal Hicator</li> <li>e. Back plunger dial indicator</li> <li>f. Deflection Gauge</li> <li>g. Special Dial Indicator</li> <li>h. Micro Indicator</li> </ul> <p>Range:                  Resolution:</p> <p>0 mm to 100 mm      1 <math>\mu</math>m                               5 <math>\mu</math>m                               10 <math>\mu</math>m</p>	<p>Technical Procedure Section BSD-08</p>	<table border="0"> <tr> <td><u>Lab</u> 1 <math>\mu</math>m 5 <math>\mu</math>m 10 <math>\mu</math>m</td> <td><u>On-site</u> 2 <math>\mu</math>m 5 <math>\mu</math>m 10 <math>\mu</math>m</td> </tr> </table>	<u>Lab</u> 1 $\mu$ m 5 $\mu$ m 10 $\mu$ m	<u>On-site</u> 2 $\mu$ m 5 $\mu$ m 10 $\mu$ m										
<u>Lab</u> 1 $\mu$ m 5 $\mu$ m 10 $\mu$ m	<u>On-site</u> 2 $\mu$ m 5 $\mu$ m 10 $\mu$ m													

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 26 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)																						
<p>9. Lever Indicating Instruments</p> <p>a. Dial test Indicator b. Pocket type dial test indicator</p> <p>Accuracy: 0 mm to 5 mm</p>	<p>Technical Procedure Section BSD -09</p>	<table border="0"> <tr> <td style="text-align: center;"><u>Lab</u></td> <td style="text-align: center;"><u>On-site</u></td> </tr> <tr> <td style="text-align: center;">1 <math>\mu</math>m</td> <td style="text-align: center;">2 <math>\mu</math>m</td> </tr> </table>	<u>Lab</u>	<u>On-site</u>	1 $\mu$ m	2 $\mu$ m																		
<u>Lab</u>	<u>On-site</u>																							
1 $\mu$ m	2 $\mu$ m																							
<p>10. Limit Gauges</p> <p>a. Plain Plug &amp; Pin Gauges Range: 0 mm to 100 mm 101 mm to 300 mm</p> <p>b. Plain Ring Gauges</p> <p>Range: 0 mm to 100 mm 101 mm to 300 mm</p> <p>c. Plain Gap Gauges Range: 0 mm to 100 mm 101 mm to 300 mm</p> <p>d. Other limit Gauges(Including Height, Depth &amp; Length) Range: 0 mm to 100 mm 101 mm to 300 mm</p> <p>e. Precision Balls Range: 0 mm to 100 mm</p>	<p>Technical Procedure Section BSD-10 &amp; BSD-11</p>	<table border="0"> <tr> <td style="text-align: center;"><u>Lab</u></td> <td style="text-align: center;"><u>On-site</u></td> </tr> <tr> <td style="text-align: center;">1 <math>\mu</math>m</td> <td style="text-align: center;">2 <math>\mu</math>m</td> </tr> <tr> <td style="text-align: center;">2 <math>\mu</math>m</td> <td style="text-align: center;">3 <math>\mu</math>m</td> </tr> <tr> <td style="text-align: center;"><u>Lab</u></td> <td style="text-align: center;"><u>On-site</u></td> </tr> <tr> <td style="text-align: center;">1 <math>\mu</math>m</td> <td style="text-align: center;">2 <math>\mu</math>m</td> </tr> <tr> <td style="text-align: center;">2 <math>\mu</math>m</td> <td style="text-align: center;">3 <math>\mu</math>m</td> </tr> <tr> <td style="text-align: center;"><u>Lab</u></td> <td style="text-align: center;"><u>On-site</u></td> </tr> <tr> <td style="text-align: center;">1 <math>\mu</math>m</td> <td style="text-align: center;">2 <math>\mu</math>m</td> </tr> <tr> <td style="text-align: center;">2 <math>\mu</math>m</td> <td style="text-align: center;">3 <math>\mu</math>m</td> </tr> <tr> <td style="text-align: center;"><u>Lab</u></td> <td style="text-align: center;"><u>On-site</u></td> </tr> <tr> <td style="text-align: center;">1 <math>\mu</math>m</td> <td style="text-align: center;">2 <math>\mu</math>m</td> </tr> </table>	<u>Lab</u>	<u>On-site</u>	1 $\mu$ m	2 $\mu$ m	2 $\mu$ m	3 $\mu$ m	<u>Lab</u>	<u>On-site</u>	1 $\mu$ m	2 $\mu$ m	2 $\mu$ m	3 $\mu$ m	<u>Lab</u>	<u>On-site</u>	1 $\mu$ m	2 $\mu$ m	2 $\mu$ m	3 $\mu$ m	<u>Lab</u>	<u>On-site</u>	1 $\mu$ m	2 $\mu$ m
<u>Lab</u>	<u>On-site</u>																							
1 $\mu$ m	2 $\mu$ m																							
2 $\mu$ m	3 $\mu$ m																							
<u>Lab</u>	<u>On-site</u>																							
1 $\mu$ m	2 $\mu$ m																							
2 $\mu$ m	3 $\mu$ m																							
<u>Lab</u>	<u>On-site</u>																							
1 $\mu$ m	2 $\mu$ m																							
2 $\mu$ m	3 $\mu$ m																							
<u>Lab</u>	<u>On-site</u>																							
1 $\mu$ m	2 $\mu$ m																							

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 27 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)	
<p>11. Line Standard</p> <p>11.1 Tapes:</p> <ul style="list-style-type: none"><li>a. Measuring Tape</li><li>b. Carpenter Tape</li><li>c. MMC tape</li><li>d. Tank Gauging Tape</li><li>e. Hook Tape</li><li>f. Circumferential Tape</li><li>g. PI Tape</li><li>h. Surveyor Tape</li></ul> <p>Range: 0 m to 50 m</p>	<p>Technical Procedure Section BSD-16</p>	<p><u>Lab</u> 1 mm</p>	<p><u>On-site</u> 1.2 mm</p>
<p>11.2 Steel Ruler</p> <p>Range: 0 m to 2000 m</p>	<p>Technical Procedure Section BSD-17</p>	<p><u>Lab</u> 0.06mm</p>	<p><u>On-site</u> 0.10 mm</p>
<p>11.3 Scales:</p> <ul style="list-style-type: none"><li>a. Coolant proof Digimatic Scale Units</li><li>b. Linear Scale</li><li>c. Glass Scale</li><li>d. Stage micrometer Scale</li><li>e. Objective Scale</li><li>f. Microscope Scale</li><li>g. Standard Scale</li></ul> <p>Range: 0 mm to 300 mm</p>	<p>Technical Procedure Section BSD-18</p>	<p><u>Lab</u> 1 <math>\mu</math>m</p>	<p><u>On-site</u> 2 <math>\mu</math>m</p>
<p>12. Protractor Standard</p> <ul style="list-style-type: none"><li>a. Digimatic Protractor</li><li>b. Dial protractor</li><li>c. Vernier Protractor</li><li>d. Bevel Protractor</li></ul>	<p>Technical Procedure Section BSD-19</p>		

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 28 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)	
<p>Range: 0 to 360 deg</p> <p>Straightness &amp; Parallelism</p> <p>13. Length Measuring Machine</p> <p>a. Horizontal Measuring System b. Universal Length Machine c. Straightness Machine d. Mic Trac Machine e. Straight Master f. Lathe Machine</p>	<p>Technical Procedure Section BSD-20</p>	<p><u>Lab</u> 5 minutes</p> <p>2 µm</p>	<p><u>On-site</u> 5 minutes</p> <p>2 µm</p>
<p>Range:                      Resolution: 0 mm to 1000 mm      0.1 µm 0 mm to 1000 mm      1 µm 0 mm to 1000 mm      5 µm 0 mm to 1000 mm      10 µm</p>		<p><u>Lab</u> 0.3 µm 1 µm 2 µm 5 µm</p>	<p><u>On-site</u> 0.5 µm 2 µm 5 µm 10 µm</p>
<p>14. Thread Plug Gauge</p> <p>Range: 0.5 mm to 300 mm</p>	<p>Technical Procedure Section BSD-22</p>	<p><u>Lab</u> 2 µm</p>	<p><u>On-site</u> 3 µm</p>
<p>15. Thread Ring Gauge</p> <p>Range: 1 mm to 100 mm</p>	<p>Technical Procedure Section BSD-23</p>	<p><u>Lab</u> 2 µm</p>	<p><u>On-site</u> 3 µm</p>
<p>16. Holtest Micrometer</p> <p>Range: 0 mm to 200 mm</p>	<p>Technical Procedure Section BSD-24</p>	<p><u>Lab</u> 2 µm</p>	<p><u>On-site</u> 3 µm</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 29 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)								
<p>17. Precision level</p> <ul style="list-style-type: none"> <li>a. Spirit Level</li> <li>b. Inclinator</li> <li>c. Precision Level</li> <li>d. Digital Level Gauge</li> <li>e. Round Level</li> <li>f. Precision Square Level</li> <li>g. Precision Flat Level</li> <li>h. Adjustable Bench Level</li> <li>i. Micro type Level</li> <li>j. Pocket Level</li> <li>k. Pendulum</li> </ul> <p>Accuracy</p> <p>Range: 0 mm to 1000 mm</p>	<p>Technical Procedure Section BSD-25</p>	<p>½ Division</p>								
<p>18. Dial Gage Tester / Calibration Tester / Dial Gage Calibrator</p> <p>Range:                      Resolution: 0 mm to 100 mm      0.1 µm 0 mm to 100 mm      1 µm 0 mm to 100 mm      10 µm</p>	<p>Technical Procedure Section BSD-26</p>	<table border="0"> <tr> <td><u>Lab</u></td> <td><u>On-site</u></td> </tr> <tr> <td>0.3 µm</td> <td>0.5 µm</td> </tr> <tr> <td>1 µm</td> <td>2 µm</td> </tr> <tr> <td>10 µm</td> <td>10 µm</td> </tr> </table>	<u>Lab</u>	<u>On-site</u>	0.3 µm	0.5 µm	1 µm	2 µm	10 µm	10 µm
<u>Lab</u>	<u>On-site</u>									
0.3 µm	0.5 µm									
1 µm	2 µm									
10 µm	10 µm									
<p>19. Lever Measuring Instruments</p> <ul style="list-style-type: none"> <li>a. Mu-Checker</li> <li>b. Inductive probe</li> </ul> <p>Range:                      Resolution: 0 mm to 10 mm      0.1 µm 0 mm to 10 mm      1 µm 0 mm to 10 mm      10 µm</p>	<p>Technical Procedure Section BSD-27</p>	<table border="0"> <tr> <td><u>Lab</u></td> <td><u>On-site</u></td> </tr> <tr> <td>0.3 µm</td> <td>0.5 µm</td> </tr> <tr> <td>1 µm</td> <td>2 µm</td> </tr> <tr> <td>10 µm</td> <td>10 µm</td> </tr> </table>	<u>Lab</u>	<u>On-site</u>	0.3 µm	0.5 µm	1 µm	2 µm	10 µm	10 µm
<u>Lab</u>	<u>On-site</u>									
0.3 µm	0.5 µm									
1 µm	2 µm									
10 µm	10 µm									

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 30 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)	
<p>20. Linear Indicating Instruments</p> <ul style="list-style-type: none"> <li>a. Linear gauge</li> <li>b. Linear Transducer</li> <li>c. Digimatic indicator</li> <li>d. Lever Gauge</li> </ul> <p>Range:                      Resolution:</p> <p>0 mm to 100 mm          0.1 <math>\mu</math>m</p> <p>0 mm to 100 mm          1 <math>\mu</math>m</p> <p>0 mm to 100 mm          10 <math>\mu</math>m</p>	<p>Technical Procedure Section BSD-28</p>	<p><u>Lab</u></p> <p>0.3 <math>\mu</math>m</p> <p>1 <math>\mu</math>m</p> <p>10 <math>\mu</math>m</p>	<p><u>On-site</u></p> <p>0.5 <math>\mu</math>m</p> <p>2 <math>\mu</math>m</p> <p>10 <math>\mu</math>m</p>
<p>21. Bore Gauge</p> <ul style="list-style-type: none"> <li>a. Length Accuracy</li> <li>b. Bore gage</li> <li>c. Bore gage (for extra small holes)</li> <li>d. Bore gage (small holes)</li> <li>e. Absolute Digimatic Bore gage</li> <li>f. Dial Bore Gage</li> <li>g. Bore Gage (for blind holes)</li> <li>h. Small Hole gage set</li> <li>i. Telescoping Gage set</li> </ul> <p>Range:</p> <p>0 mm to 600 mm</p>	<p>Technical Procedure Section BSD-29</p>	<p><u>Lab</u></p> <p>1 <math>\mu</math>m</p>	<p><u>On-site</u></p> <p>2 <math>\mu</math>m</p>
<p>22. Feeler Gauge</p> <p>Range:</p> <p>Up to 10 mm</p>	<p>Technical Procedure Section BSD-31</p>	<p><u>Lab</u></p> <p>1 <math>\mu</math>m</p>	<p><u>On-site</u></p> <p>2 <math>\mu</math>m</p>
<p>23. Linear height Gauge</p> <p>Range:</p> <p>0 mm to 600 mm</p> <p>0 mm to 1000 mm</p>	<p>Technical Procedure Section BSD-34</p>	<p><u>Lab</u></p> <p>2 <math>\mu</math>m</p> <p>3 <math>\mu</math>m</p>	<p><u>On-site</u></p> <p>3 <math>\mu</math>m</p> <p>5 <math>\mu</math>m</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 31 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)	
<p>24. Micrometer Head</p> <p>Range: 0 mm to 25 mm 0 mm to 50 mm 0 mm to 100 mm</p>	<p>Technical Procedure Section BSD-35</p>	<p><u>Lab</u> 1 µm 1 µm 1 µm <u>Lab</u></p>	<p><u>On-site</u> 2 µm 2 µm 2 µm <u>On-site</u></p>
<p>25. Depth Micro checker/Caliper checker</p> <p>Flatness :</p> <p>Parallelism: 0 mm to 1000 mm</p> <p>Accuracy: 0 mm to 1000 mm</p>	<p>Technical Procedure Section BSD-37</p>	<p><u>Lab</u> 2 µm  <u>Lab</u> 2 µm  <u>Lab</u> 1 µm</p>	<p><u>On-site</u> 3 µm  <u>On-site</u> 3 µm  <u>On-site</u> 2 µm</p>
<p>26. Measuring Toolmaker</p> <p>Toolmaker Microscope Microscope Smart Scope Vision unit Video Measuring System</p> <p>Range: 0 mm to 250 mm-Y axis 0 mm to 250 mm-Y &amp; Z axis 0 mm to 360 ° Magnification up to 100x</p>	<p>Technical Procedure Section BSD-39</p>	<p><u>Lab</u> 0.3 µm 1 µm 10 µm 0.1 %</p>	<p><u>On-site</u> 0.5 µm 2 µm 10 µm 0.2 %</p>
<p>27. Measuring Profile</p> <p>Profile Projector Optical Comparator</p> <p>Range : 0 ~ 250 mm-Y axis 0 ~ 250 mm-Y &amp; Z axis 0 ~ 360 ° Magnification up to 100x</p>	<p>Technical Procedure Section BSD-40</p>	<p><u>Lab</u> 0.3 µm 1 µm 10 µm 0.1 %</p>	<p><u>On-site</u> 0.5 µm 2 µm 10 µm 0.2 %</p>

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 32 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)						
<p>28. Thickness Standard</p> <p>28.1</p> <ul style="list-style-type: none"><li>a. Dial thickness Gauge</li><li>b. Digimatic Thickness Gauge</li><li>c. Pocket Thickness Gauge</li><li>d. Surface Profile Gauge</li><li>e. Light Weight Dial</li><li>f. Thickness Gauge</li><li>g. Tube Thickness</li><li>h. Ultra sonic thickness gauge</li></ul> <p>Range :</p> <p>0 mm to 10 mm</p> <p>0 mm to 200 mm</p>	<p>Technical Procedure Section BSD-42</p>	<table><tr><td><u>Lab</u></td><td><u>On-site</u></td></tr><tr><td>2 µm</td><td>3 µm</td></tr><tr><td>2 µm</td><td>3 µm</td></tr></table>	<u>Lab</u>	<u>On-site</u>	2 µm	3 µm	2 µm	3 µm
<u>Lab</u>	<u>On-site</u>							
2 µm	3 µm							
2 µm	3 µm							
<p>28.2</p> <ul style="list-style-type: none"><li>a. Coating Thickness foils</li></ul> <p>Range:</p> <p>0 mm to 10 mm</p>	<p>Technical Procedure Section BSD-31</p>	<table><tr><td><u>Lab</u></td><td><u>On-site</u></td></tr><tr><td>1 µm</td><td>2 µm</td></tr></table>	<u>Lab</u>	<u>On-site</u>	1 µm	2 µm		
<u>Lab</u>	<u>On-site</u>							
1 µm	2 µm							
<p>29. Geometric Measurement</p> <ul style="list-style-type: none"><li>a. Radius Gauge</li><li>b. Pitch Gauge</li><li>c. Taper Gauge</li><li>d. Jigs</li><li>e. Step Block</li><li>f. Fixtures</li><li>g. Angle Gauge</li><li>h. Limit Gauges</li><li>i. Height Setting Gauge</li><li>j. Depth gauge</li></ul>	<p>Technical Procedure Section BSD-51</p>							



# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 33 of 34

MEASURED QUANTITIES / INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (*)	
Flatness:  Parallelism: 0 to 300 mm  Straightness: 0 mm to 1500 mm  Angle: 0 deg to 360 deg  Distance/Length: 0 mm to 1000 mm  Radius: 0 mm to 100 mm  Diameter: 0 mm to 200 mm  30. Test Sieve  a. Metal Cloth wire size Range: Up to 15 mm  b. Perforated Metal Plate Range: Up to 125 mm	Technical Procedure Section BSD-52	<u>Lab</u> 2 µm  <u>Lab</u> 2 µm  <u>Lab</u> 2 µm  <u>Lab</u> 5 minutes  <u>Lab</u> 4 µm  <u>Lab</u> 4 µm  <u>Lab</u> 3 µm   <u>Lab</u> 5 µm  <u>Lab</u> 8 µm	<u>On-site</u> 3 µm  <u>On-site</u> 3 µm  <u>On-site</u> 3 µm  <u>On-site</u> 5 minutes  <u>On-site</u> 5 µm  <u>On-site</u> 5 µm  <u>On-site</u> 5 µm   <u>On-site</u> 8 µm  <u>On-site</u> 10 µm

\* A reported uncertainty will be that for the instrument itself during calibration plus the appropriate measurement capability of the laboratory. The uncertainties are based on an estimated confidence probability of approximately 95% unless otherwise stated.

# Schedule



Certificate No. : LA-2006-0358-C

Issue No. : 9

Date : 18 October 2013

Page : 34 of 34

## Approved Signatories

- Mr B Balamurga Redhy ) For all items
- Mr Sekar Karuppiah ) For all items B and C (except Liquid-in-glass thermometer)
- Mr R Gnanasekar ) Only for item B1 to B9

## Note :

This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025:2005. A laboratory's fulfilment of the requirements of ISO/IEC 17025:2005 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibration results. The **management system requirements** in ISO/IEC 17025:2005 (Section 4) are written in language relevant to laboratory operations and meet the principles of ISO 9001:2008 **Quality Management Systems — Requirements** and are aligned with its pertinent requirements.